



QuickVue+  
hCG Combo TEST

## CLIA Complexity: Waived-Urine / Moderate-Serum

For *in vitro* diagnostic use.

**Rx ONLY**

A symbols glossary can be found at [quidel.com/glossary](http://quidel.com/glossary).



The QuickVue+ hCG Combo Test is a one-step immunoassay intended for the qualitative detection of human Chorionic Gonadotropin (hCG) in serum or urine for the early detection of pregnancy. The test is intended for use by healthcare professionals.

## SUMMARY AND EXPLANATION

Human chorionic gonadotropin is a hormone normally produced by the placenta. Since hCG is present in the serum and urine of pregnant women, it is an excellent marker for confirming pregnancy.

The QuickVue+ hCG Combo Test is a lateral-flow test using a monoclonal antibody specific to the beta subunit of hCG to accurately detect hCG as early as 2 or 3 days before the expected menses.

## PRINCIPLE OF THE TEST

To perform the test, a serum or urine sample is collected and added to the Reaction Unit. If the sample contains hCG, a pink vertical line forms in the Read Result Window. This pink vertical line, together with the pre-printed blue horizontal line, form a plus sign (+) to indicate a positive result. If hCG is not present in the sample, the Read Result Window shows only the pre-printed blue horizontal line, forming a minus sign (–) to indicate a negative result.

As the sample continues to move through the test, a bar in the Control Window becomes blue. Blue color in the Control Window indicates that the test is functionally active and is also evidence that the test has been performed correctly.

## REAGENTS AND MATERIALS SUPPLIED

Each QuickVue+ hCG Combo Test kit contains enough materials for 30 tests (Cat. #00178) or 90 tests (Cat. #00179).

- Reaction Unit (30 or 90)
  - ▶ Contains a murine monoclonal antibody and a caprine polyclonal antibody to hCG
- Disposable Dropper (30 or 90)
- Package Insert (1)
- Procedure Card (1)

## MATERIALS REQUIRED BUT NOT PROVIDED

- Timer or watch that measures minutes and seconds
- Specimen collection containers

## WARNINGS AND PRECAUTIONS

- For *in vitro* diagnostic use
- Do not use kit contents after the expiration date printed on the outside of the kit.
- Use appropriate precautions in the collection, storage, handling and disposal of patient samples and used kit contents.
- Use of Nitrile or Latex gloves is recommended when handling patient samples.<sup>1</sup>
- The Reaction Unit must remain sealed in the foil pouch just prior to use.
- To obtain accurate results, you must follow the Package Insert instructions.
- Testing should be performed in an area with adequate ventilation.
- Dispose of containers and unused contents in accordance with Federal, State and Local regulatory requirements.
- Wear suitable protective clothing, gloves, and eye/face protection when handling the contents of this kit.
- Wash hands thoroughly after handling.
- For additional information on hazard symbols, safety, handling and disposal of the components within this kit, please refer to the Safety Data Sheet (SDS) located at [quidel.com](http://quidel.com).

## KIT STORAGE AND STABILITY

Store kit at room temperature, 59°F to 86°F (15°C to 30°C), out of direct sunlight. Do not freeze.

## SPECIMEN COLLECTION AND STORAGE

### *Urine*

Collect urine specimens in a clean container. Urine collected anytime during the day can be used. For optimal results, it is best to test the first urine voided in the morning because it contains the greatest concentration of hCG. Samples can be stored for 8 hours at room temperature (59°F to 86°F; 15°C to 30°C) or up to 72 hours refrigerated (36°F to 46°F; 2°C to 8°C). DO NOT freeze the urine sample.

### *Serum*

No special patient preparation is necessary. A whole blood specimen should be obtained by standard medical procedures. After clotting has occurred, the separated serum should be used for testing.

Serum specimens may be stored refrigerated (2°C to 8°C) for up to 48 hours prior to assay. If testing will be delayed for more than 48 hours, the sample may be frozen once at -20°C or below. If frozen, mix after thawing. Do not re-freeze. Do not chemically modify the serum in any way.

## QUALITY CONTROL

### *External Quality Control*

Positive and negative controls can be run with each shipment of a new kit lot number and as otherwise required by your laboratory's quality assurance plan. The hCG Control Set — Serum (Cat. #00281) or the hCG Control Set — Urine (Cat. #00272) can be used for this purpose. The use of other hCG controls may be incompatible with the assay.

- **External Positive Control:** Process the control as you would a patient sample. A positive signal is indicated by a pink and blue plus sign (+) in the Read Result Window along with a blue procedural Control Line in the Control Window.

- **External Negative Control:** Process the control as you would a patient specimen. A negative signal is indicated by a blue minus sign (–) in the Read Result Window along with a blue procedural Control Line in the Control Window.

### Internal Control Features

The QuickVue+ hCG Combo Test contains built-in control features. The manufacturer’s recommendation for daily quality control is to document these controls for the first sample tested each day. Quality Control log sheets are available from Technical Support. No additional external quality control is required.

- **Internal Positive Procedural Control:** A blue line in the Control Window is considered an internal positive procedural control. If the test has been performed correctly and the Reaction Unit is working properly, this indicator will appear.
- **Internal Negative Procedural Control:** A clear background in the Read Result Window is considered an internal negative procedural control. If the test has been performed correctly and the Reaction Unit is working properly, the background will clear to give a discernable result.

If the controls do not perform as expected, do not use the test results. Repeat the test or contact Technical Support.

### PROCEDURAL NOTES

- DO NOT remove the Reaction Unit from the foil pouch until you are ready to perform the test.
- Use a new Disposable Dropper for each sample to avoid cross-contamination.

### TEST PROCEDURE

- Remove the Reaction Unit from the pouch.
- Draw serum or urine into the Disposable Dropper and dispense **4 drops** into the Add Sample well.
- Shortly after the sample is added, a pink-to-purple color will be seen moving across the Reaction Unit’s windows. The Read Result Window contains a pre-printed horizontal blue line on the membrane.
- **Read result at 3 minutes for urine and 5 minutes for serum.**  
**Note:** Some positive results may appear sooner.

### INTERPRETATION OF RESULTS

Refer to Procedure Card for interpretation of test results.

#### Positive Result:

The sample contains a detectable amount of hCG when you see: A pink and blue plus sign (+) in the large square Read Result Window, along with a blue line in the small square Control Window.



**Note:** Any shade of a pink vertical line in the Read Result Window should be interpreted as a positive result.

#### Negative Result:

The sample does not contain detectable amounts of hCG when you see: A blue minus sign (–) in the large square Read Result Window, along with a blue line in the small square Control Window.



---

### Invalid Result:

The result is invalid if: No blue line appears in the small square Control Window; or background color in the large square Read Result Window interferes with test interpretation. In the case of an invalid result, a new patient specimen should be tested using a new QuickVue+ hCG Combo Test or contact Technical Support.



If a negative result is obtained, but pregnancy is suspected, another sample should be collected after 48-72 hours and tested.

---

## LIMITATIONS

- The contents of this kit are for use in the **qualitative** detection of hCG in serum or urine.
- Test results must always be evaluated with other data available to the physician.
- A normal pregnancy cannot be distinguished from an ectopic pregnancy based on hCG levels alone.<sup>2</sup>
- Very low levels of hCG are present in serum and in urine shortly after implantation. Positive test results from very early pregnancy may later prove negative due to natural termination of pregnancy. This is estimated to occur in up to 50% of all conceptions.<sup>3</sup> If a very low, faint positive serum result is obtained, another sample should be obtained in 48 hours and retested. If waiting 48 hours is not medically advisable, the test result should be confirmed with a quantitative hCG test.
- Patients with trophoblastic and nontrophoblastic disease may have elevated hCG levels, therefore, the possibility of hCG secreting neoplasms should be eliminated prior to the diagnosis of pregnancy.<sup>4</sup>
- If a urine sample is too dilute, it may not contain a representative urinary hCG concentration. If a negative result is obtained and pregnancy is still suspected, a first morning sample should be obtained and tested.

## EXPECTED VALUES

The sensitivity of QuickVue+ hCG Combo Test is 10 mIU/mL for serum or 20 mIU/mL for urine (WHO 3rd IS 75/537). In normal pregnancy, hCG levels in urine can reach 25 mIU/mL as early as 7-10 days post conception, and continue to increase exponentially to reach a maximum concentration in excess of 200,000 mIU/mL at the end of the first trimester.<sup>5</sup>

## PERFORMANCE CHARACTERISTICS

### Clinical Sensitivity, Specificity, and Accuracy

The clinical sensitivity, specificity and accuracy of the QuickVue+ hCG Combo Test was determined by evaluating Four hundred ninety-seven (497) urine samples and One thousand five hundred twenty-eight (1528) serum samples obtained from women presenting for pregnancy testing. Testing was performed by clinic personnel. Results were compared to results obtained with the Abbott TestPack® Plus hCG-Combo. A commercially available quantitative method was used to resolve any discrepant results.

The results are summarized below. Of the Four hundred ninety-seven (497) urine samples tested, the QuickVue+ hCG Combo Test yielded an accuracy of > 99%.

Similarly, of the One thousand five hundred twenty-eight (1528) serum samples tested, the QuickVue+ hCG Combo Test yielded an accuracy of > 99%.

		Serum Correlation TestPack Plus hCG Combo	
		+	-
QuickVue+ hCG Combo Test	+	510	6
	-	0	1012
Sensitivity:		510/510 = > 99%	
Specificity:		1012/1018 = >99%	
Accuracy:		1522/1528 = >99%	

		Urine Correlation TestPack Plus hCG Combo	
		+	-
QuickVue+ hCG Combo Test	+	269	0
	-	0	228
Sensitivity:		269/269 = >99%	
Specificity:		228/228 = >99%	
Accuracy:		497/497 = >99%	

## PHYSICIAN'S OFFICE LABORATORY (POL) STUDIES

An evaluation of the QuickVue+ hCG Combo Test was conducted at three Physicians' Offices using a panel of coded specimens. Testing was performed by physician's office personnel with diverse educational backgrounds and work experience at different locations. The proficiency panel contained negative, low positive and moderate positive samples. Each level was tested in replicates of five at each site over a period of 3 days.

The results at each site agreed 100% with the expected results. No significant differences were observed within run, between runs, or between sites.

### Cross-Reactivity

hTSH, hLH, and hFSH were tested in the QuickVue+ hCG Combo Test at levels ranging from 1000  $\mu$ IU/mL to 1000 mIU/mL and did not affect the expected results.

### Interference Testing

The following chemical and biological compounds were tested in the QuickVue+ hCG Combo Test and did not affect the expected results:

Urine Analytes	Concentration
Albumin (serum)	2000 mg/dL
Bilirubin	1000 $\mu$ g/dL
Hemoglobin	1000 $\mu$ g/dL
Glucose	2000 mg/dL
Urine pH	5-9
Hormones	Concentration
LH	300 mIU/mL
FSH	1000 mIU/mL
TSH	1000 mIU/mL
Estriol 17-beta	1400 $\mu$ g/dL
Pregnanediol	1500 $\mu$ g/dL
Bacteria	Concentration
<i>E. coli</i>	10 <sup>8</sup> CFU/mL
Group B Streptococcus	2.5 x 10 <sup>7</sup> CFU/mL
<i>Chlamydia trachomatis</i>	10 <sup>7</sup> IFU/mL
Chemical Analytes	Concentration
Acetaminophen	20 mg/dL
Acetoacetic Acid	2000 mg/dL
Ascorbic Acid	20 mg/dL

β-Hydroxybutyrate	2000 mg/dL
Biotin	< 714 ng/mL
Caffeine	20 mg/dL
Clomiphene	100 mg/dL
Gentisic Acid	20 mg/dL
Salicylic Acid	20 mg/dL
EDTA	80 mg/dL
Cannabinol	10 mg/dL
Cocaine	10 mg/dL
Codeine	10 mg/dL
Heroin	1 mg/dL
Methadone	10 mg/dL
Methamphetamine	10 mg/dL
Methanol	10.0%
Ephedrine	20 mg/dL
Phenylpropanolamine	20 mg/dL
Theophylline	20 mg/dL
Acetylsalicylic Acid	20 mg/dL
Benzoyllecgonine (cocaine metabolite)	10 mg/dL
Ethanol	1.0%
DMSO	5.0%
Uric Acid	20 mg/dL
Heparin	2800 units/dL

## ASSISTANCE

If you have any questions regarding the use of this product, please contact Quidel Technical Support at 1.800.874.1517 (in the U.S.) or [technicalsupport@quidel.com](mailto:technicalsupport@quidel.com). If outside the U.S., further information can be obtained from your distributor, or directly from Quidel at one of the numbers listed below. Reference [quidel.com](http://quidel.com) to see more options for Support.

Country	Phone	E Mail Address
Europe, Middle East and Africa	+353 (91) 412 474 (main) 0 1800 200441 (toll free)	<a href="mailto:emeatechnicalsupport@quidel.com">emeatechnicalsupport@quidel.com</a>
Austria	+43 316 231239	
France	0 (805) 371674	
Germany	+49 (0) 7154 1593912	
Netherlands	0 800 0224198	
Switzerland	0 800 554864	
United Kingdom	0 800 3688248	
Italy	+39 (800) 620 549	
North America, Asia-Pacific, Latin America	858.552.1100	<a href="mailto:technicalsupport@quidel.com">technicalsupport@quidel.com</a>
Canada	437.266.1704 (main) 888.415.8764 (toll free)	<a href="mailto:technicalsupport@quidel.com">technicalsupport@quidel.com</a>
China	0400 920 9366 or +86 021 3217 8300	<a href="mailto:chinatechnicalservice@quidel.com">chinatechnicalservice@quidel.com</a>

## REFERENCES

1. Biosafety in Microbiological and Biomedical Laboratories, 5th Edition. U.S. Department of Health and Human Services, CDC, NIH, Washington, DC (2009).

2. Razor J.L. and Braunstein G.D. Obstet. Gynecol., 1977 50, 553-558.
3. Edmonds D.K., Lindsay K.S., Miller J.F., Williamson E. and Wood R.J. Fertility and Sterility, 1982 38, 447-453.
4. Braunstein G.D., Vaitukaitis J.L., Carbone P.P., and Ross G.T. Ann. Intern. Med., 1973 78, 39-45.
5. Lenton E.A., Neal L.M., and Sulaiman R. Fertility and Sterility, 1982 37, 773-778.

REF

00178 – QuickVue+ hCG Combo 30 Test Kit  
00179 – QuickVue+ hCG Combo 90 Test Kit  
00272 – hCG Control Set – Urine  
00281 – hCG Control Set – Serum

IVD



EC REP

MDSS GmbH  
Schiffgraben 41  
30175 Hannover,  
Germany



**Quidel Corporation**  
10165 McKellar Court  
San Diego, CA 92121 USA  
[quidel.com](http://quidel.com)



0436409EN00 (07/21)

---

**REF**

Catalogue number



CE mark of conformity

---

**EC REP**

Authorized Representative  
in the European Community

**LOT**

Batch code

---



Use by



Manufacturer

---



Temperature limitation



Intended use

---

**R<sub>x</sub> ONLY**

Prescription use only



Consult instructions for use

---

**IVD**

For *In Vitro* diagnostic use



Contains sufficient for XX determinations

---

**CONT**

Contents/Contains

---