

HUNTLEIGH MD2/SD2

Anwendungshinweise

Kullanım Talimatları

Brugsvejledning

Instrucciones de uso

χρήσης

使用方

Mode d'emploi

Bruksanvisning

Gebruiksaanwijzing

aanwijzing

INSTRUCTIONS FOR USE

Bruksan

talimatları

使用方法

Käyttöohjeet

Instruções de Utilização

Istruzioni per l'uso

Anwendungshinweise

Οδηγίες χρήσης

Anwendungshinweise

HIGH SENSITIVITY POCKET DOPPLERS

Table of Contents

Doppler Measurement Sites and Recommended Probes 3

1. Safety	4
1.1 Warnings	4
1.2 Patient Applied Parts	5
1.3 Acoustic Safety	5
2. Introduction	7
2.1 Unpacking / Preliminary Checks	7
2.2 Battery Insertion / Replacement	7
2.3 Product Controls	8
2.4 Product Labelling	9
3. Operation	10
3.1 Vascular Mode	10
3.2 Obstetric Mode	12
3.3 After Use	14
4. Care and Cleaning	15
4.1 General Care	15
4.2 General Cleaning and Disinfecting	16
4.3 Cleaning and Disinfecting Patient Applied Parts	16
4.4 Maintenance and Repair	17
5. Specifications	18
5.1 Equipment Classification	18
5.2 Standards Compliance	18
5.3 FHR Performance	18
5.4 Waveform Outputs (MD2 Only)	19
5.5 General	19
5.6 Environmental	19
6. End of Life Disposal.....	20
7. Warranty	21
8. Service.....	23

Doppler Measurement Sites and Recommended Probes

1 Jugular Vein
VP4HS, VP5HS

2 Subclavian Vein
VP4HS, VP5HS

3 Fetus
OP2HS, OP3HS

4 Femoral Vein
VP4HS, VP5HS

5 Great Saphenous Vein
VP5HS, VP8HS, EZ8

6 Small Saphenous Vein
VP8HS, EZ8

7 Posterior Tibial Vein
VP8HS, VP10HS, EZ8

8 Posterior Tibial Artery
VP8HS, EZ8

9 Vertebral Artery
VP4HS, VP5HS

10 Carotid Artery
VP5HS, VP8HS, EZ8

11 Subclavian Artery
VP4HS, VP5HS

12 Brachial Artery
VP8HS, EZ8

13 Ulnar Artery
VP8HS, EZ8

14 Radial Artery
VP8HS, EZ8

15 Femoral Artery
VP4HS, VP5HS


16 Digital Artery
VP8HS, VP10HS, EZ8

17 Penis
VP10HS, EZ8

18 Popliteal Artery
VP5HS

19 Metatarsal Artery
VP8HS, VP10HS, EZ8

20 Dorsalis Pedis Artery
VP8HS, VP10HS, EZ8



1. Safety



Before using this equipment, please study this manual carefully and familiarize yourself with the controls, display features and operation. Ensure that each user fully understands the safety and operation of the unit, as mis-use may cause harm to the user or patient, or damage to the product.



We recommend that exposure to ultrasound should be kept As Low As Reasonably Achievable - (ALARA guidelines). This is considered to be good practice and should be observed at all times.



Federal law restricts this device to sale by or on the order of a licensed practitioner

Please keep these Instructions for Use to hand for future reference.



Attention, consult this manual. Refer to safety section.



Attention, consult accompanying documents / Instructions for Use

1.1 Warnings



Do not use in the presence of flammable gases such as anesthetic agents.



Do not use in the sterile field unless additional barrier precautions are taken.



Do Not :

- ***immerse in any liquid,***
- ***use solvent cleaner,***
- ***use high temperature sterilizing processes (such as autoclaving),***
- ***use E-beam or gamma radiation sterilization.***



Do not use on the eye or scrotum.



Do not dispose of batteries in fire as this can cause them to explode.



Do not attempt to recharge normal dry-cell batteries. They may leak, cause a fire or even explode.



This product contains sensitive electronics, therefore, strong radio frequency fields could possibly interfere with it. This will be indicated by unusual sounds from the loudspeaker. We recommend that the source of interference is identified and eliminated.



Any equipment connected to RS232 interface must be compliant with IEC60601-1:2005.



Connect headphones only to the headphone socket.



Dopplex Dopplers are screening tools to aid the healthcare professional and should not be used in place of normal vascular or fetal monitoring. If there is doubt as to vascularity or fetal well-being after using the unit, further investigations should be undertaken immediately using alternative techniques.

1.2 Patient Applied Parts

As defined in IEC60601-1:2005, the patient applied parts of the Dopplex Dopplers are the ultrasound probes.

1.3 Acoustic Safety

Continuous wave Doppler ultrasound instruments such as the MD2/SD2 have been used extensively for medical diagnosis in the United States for over 25 years. Throughout this period, there have been no reports of adverse effects to patients or instrument operators at the acoustic intensities recommended for diagnostic use. Despite this highly favorable safety experience, available data are not conclusive and the possibility remains that unwanted biological effects might be identified in the future. Authorities therefore recommend that ultrasound procedures be performed in accordance with the "ALARA" principle, which states that the energy delivered to the patient should always be kept As Low As Reasonably Achievable. With the MD2/SD2, the transmitted acoustic power is fixed and cannot be adjusted by the operator. Therefore, the user can best observe the ALARA principle by ensuring that each examination is medically indicated and by limiting the duration of the study to the extent appropriate for the clinical objectives.

Acoustic intensity data ($I_{SPTA,3}$) for probes available for use with the MD/SD2 are summarized in the following table. The values cited are based on measurements in water using a calibrated hydrophone and are stated as the estimated derated intensities. The derated intensity constitutes the most biologically relevant parameter available since true determinations of actual absorbed dose in tissue would require invasive measurement techniques. The derated intensity is therefore calculated mathematically using a derating factor consisting of a constant (the assumed attenuation coefficient) and allowing for the frequency of the probe and the distance from the probe face to the hydrophone.

The calculated derated intensity values for the MD/SD2 compare very favorably with previously reported acoustic safety data for Doppler ultrasound instruments and are appropriate for all clinical applications recommended in this manual. As the operating mode of the Doppler range of probes is continuous wave, I_{SPPA} figures are not applicable.

Acoustic Output Table, Track1, Non-Auto-Scanning Mode						
Model	Max Value $I_{SPTA,3}$	Wo	f_c	Z_{sp}	A-6, (Z_{sp})	EBD
OP2HS	55	41	2.0	2.5	1.1	1.2 x 2.5
OP3HS	55	32	3.0	2.0	1.2	1.2 x 2.5
VP4HS	92	7.5	4.0	0.8	0.14	0.365 x 0.8
VP5HS	92	8.2	5.0	0.8	0.12	0.365 x 0.8
VP8HS	92	4.0	8.0	0.48	0.026	0.215 x 0.5
EZ8	92	14.3	8.0	0.67	0.064	0.635 x 0.22
VP10HS	92	1.4	10.0	0.48	0.022	0.215 x 0.5

NOTES

- Measurement uncertainty: varies with probe and measurement
 Random - typically $\pm 20\%$ (max. $\pm 32\%$)
 Systematic - typically $\pm 6.5\%$ (max. $\pm 8\%$)


Definition of Terms

- $I_{SPTA,3}$ is the derated spatial-peak, temporal-average intensity (milliwatts per square centimeter)
- Wo** is the ultrasonic power (milliwatts)
- f_c is the center frequency (MegaHertz)
- Z_{sp} is the axial distance used to calculate the derated intensity (centimeters)
- A-6 (Z_{sp})** is $(\pi / 4) \times (X-6 \times Y-6)$ where X-6, Y-6 are respectively the in-plane (azimuthal) and out-of-plane (elevational) -6dB dimensions in the X-Y plane where Z_{sp} is found (centimeters)
- EBD** are the entrance beam dimensions for the azimuthal and elevational planes (centimeters)

2. Introduction

2.1 Unpacking / Preliminary Checks

Contents




Delivery Inspection

Huntleigh takes every precaution to ensure that goods reach you in perfect condition. However, accidental damage can occur in transit and storage. For this reason we recommend that a thorough visual inspection is made immediately the unit is received. Should any damage be evident or any parts missing, ensure that Huntleigh or your distributor is informed at once.

Storage


Should the unit not be required for immediate use, it should be re-sealed into its original packing after carrying out the initial delivery inspection, and stored under covered conditions at a temperature between +14°F to +104°F (-10°C to +40°C), and relative humidity of 10% to 93% non-condensing.

2.2 Battery Insertion / Replacement




Note: Remove the battery if the unit is not likely to be used for some time.


















2.3 Product Controls



	MD2	SD2	
1	•	•	Headphone Socket
2	•		Waveform Socket
	•		RS232 Port
3	•	•	LCD Panel
4	•	•	Loud-speaker
5	•	•	On/Off Button
6	•	•	Probe Holder
7	•	•	Volume Control
8	•		Start/Stop Button
	•		Cal Button
9	•		Mode Button
		•	Gain Button
10	•	•	Battery Compartment
11	•	•	Pocket Clip



2.4 Product Labelling

	Applied parts (ultrasound probes) are type B according to the definitions in IEC60601-1:1988		
	Power On/Off		
	This symbol signifies that this product, including its accessories and consumables is subject to the WEEE (Waste Electrical and Electronic Equipment) regulations and should be disposed of responsibly in accordance with local procedures.		
	Attention, consult this manual. Refer to safety section.		
	Attention, consult accompanying documents / Instructions for Use		
	This symbol signifies that this product complies with the essential requirements of the Medical Devices Directive 93/42/EEC as amended by 2007/47/EC		
Rx only	Caution: Federal law restricts this device to sale by or on the order of a licensed practitioner.		
	Alignment mark		RS232 Interface
	Volume		Headphone Socket
	Temperature Limitations	"MAX 93% RH"	Limits of Relative Humidity
	Serial Number		Reference Number
	Keep Dry		Do not use hook
	Fragile		Cardboard packaging can be recycled.

3. Operation



Refer to diagram on page 3 for Doppler Measuring sites and Recommended Probes.

To connect the probe, align the arrow on the connector with the slot on the probe and push firmly.

To disconnect the probe, pull the connector sharply. **DO NOT** pull the cable.

Note: During use, an automatic noise reduction feature operates on low level signals to improve sound quality.

Coupling Gel

Use water based ultrasound gel **ONLY**.

3.1 Vascular Mode

The Multi Dopplex II/Super Dopplex II Dopplers (MD2/SD2) will select vascular mode when a vascular probe is connected to the control unit.

Vascular Probes

Five probes are available for vascular examinations:

VP4HS	4MHz \pm1% for deep lying vessels
VP5HS	5MHz \pm1% for deep lying vessels and oedematous limbs
VP8HS	8MHz \pm1% for peripheral vessels
VP10HS	10MHz \pm1% for specialist superficial applications.
EZ8	8MHZ \pm1% "Widebeam" for peripheral vessels.

In this mode, bi-directional blood flow rate and direction are indicated on bar graphs (4 levels in each direction) and blood flow is audible in the loudspeaker. Probe frequency is displayed together with the bar graphs.

Clinical Use

Apply a liberal amount of gel on the site to be examined. Place the probe at 45° to the skin surface over the vessel to be examined. Adjust the position of the probe to obtain the loudest audio signal. High pitched pulsatile sounds are emitted from arteries while veins emit a non-pulsatile sound similar to a rushing wind.

For best results, keep the probe as still as possible once the optimum position has been found. Adjust the audio volume as required.


Waveform Recording (MD2 only)

Separated waveform outputs are provided in analogue and digital formats. Analogue signals are provided for connection to a single channel chart recorder giving a combined bi-directional waveform, or to a dual channel recorder for separate forward and reverse flow waveforms. (Pin-out details on request).

Digital signals are provided via the RS232 port for printing separate waveforms on a Dopplex Printa II or for communicating with Dopplex Reporter software package.

Cal Function

The baseline and sensitivity of the chart recorder can be set up using the Cal function. This generates a zero velocity baseline and a sequence of bi-directional pulses as shown below:



Gain Control

To cope with the wide variety of signals detectable using your MD2/SD2, a gain control allows you to optimise the bar-graph display and increase the height of the waveforms.

Using the Mode button on the MD2 (Gain button for SD2), gain can be set to x1, x2, x4, x8.

Connection to Printa II™ (MD2 Only)

Hard copy printing is automatically selected when the plug of the interface buffer box is inserted into the RS232 socket on the top panel of the MD2. Printing is then initiated by using the Start/Stop button.

Connection to Software (MD2 Only)

The MD2 can be connected to the Reporter Software Package, via the RS232 interface. Dopplex Reporter software package is available as an accessory. Communicating is then initiated by using the Start/Stop button.

3.2 Obstetric Mode



IMPORTANT! READ THIS SECTION BEFORE USING YOUR Multi Dopplex.



Although the Multi Dopplex will calculate the rate as early in pregnancy as fetal pulse signals of good quality can be elicited, generally from about the 9th or 10th week after the last menstrual period, fetal heart rhythm may not be sufficiently stable to permit accurate rate computation until well after the first trimester.



Similarly, no attempt to apply non-stress test (NST) interpretation criteria should be made until after the 25th week of gestation, when neurological development is usually sufficient to obtain a reactive pattern. To be considered reactive, two accelerations at least 15 bpm in amplitude and at least 15 seconds in duration should occur in a 10 minute period.



If the fetus is found to be non-reactive using the Multi Dopplex hand-held sampling technique, consideration should be given to repeating the test using a conventional fetal monitor.



The Multi Dopplex is suitable for intermittent auscultation during the intrapartum period, but not recommended for long term intrapartum monitoring, where Cardiotocographic equipment, including contractions measurement capability should be employed.

Obstetric mode is automatically selected when an obstetric probe (OP2HS/OP3HS) is connected.

Obstetric Probes

Two probes are available for obstetric examinations:

OP2HS	2MHz ±1%
OP3HS	3MHz ±1%

In this mode, MD2 provides fetal heart rate (FHR) display with 3 operating modes, and outputs FHR for printing on the Printa II. The SD2 in obstetric mode provides audio only.

Clinical Use

Apply a liberal amount of gel to the abdomen. Place the faceplate of the probe flat against the abdomen above the symphysis pubis. Adjust the probe to obtain an optimum audio signal ideally by angling the probe around. Avoid sliding it over the skin.

In early pregnancy a full bladder may improve sound detection. In later pregnancy the best signals are generally located higher on the abdomen. The fetal heart sounds like a galloping horse at approximately twice the maternal rate. A wind-like sound is heard from the placenta.

Standard Mode - MD2 Only



In this mode the FHR, averaged over 4-heart beats, is displayed on the 3-digit readout. The LCD displays an outline heart symbol.

Smoothed Mode - MD2 Only



This mode is used to obtain more stable heart rate readings. In this mode, FHR is averaged over 8 beats. The LCD displays a solid heart symbol.

Manual Mode - MD2 Only



This mode is used when a fetal heart beat is audible in the loudspeaker or headphones but, due to noise or a low signal level, the MD2 cannot reliably calculate the heart rate. In this mode, the heart rate can be manually counted over a period of 10 audible heart beats (see below). The MD2 will automatically calculate and display the derived FHR on the LCD. The LCD displays a clock symbol.

Mode Selection



Press the Mode button to select mode.

Use of Manual Mode

1. Press and hold Start/Stop button and immediately count the audible heart beats, counting the first beat as the button is pressed. The LCD displays the flashing clock symbol and the FHR reading is shown as three dashes.
2. Release the Start/Stop button immediately on the count of 10 (i.e. After nine beat intervals). The MD2 will automatically calculate the derived FHR averaged over the 10 beat period and display the result. This rate value is retained until the measurement is repeated or the unit is switched off. If the button is held for a period less than about 3 seconds the display will clear the previous rate value and reset.

Connection to Printa II™ (MD2 Only)

Hard copy printing is automatically selected when the plug of the interface buffer box is inserted into the RS232 socket on the top panel of the MD2. Printing is then initiated by using the Start/Stop button.

3.3 After Use

1. Press and release the On/Off button. If you forget to switch the unit off, it will automatically shut-off after 3 minutes.
2. Refer to the cleaning section before storing or using the unit on another patient.
3. Store unit together with probe and accessories in the soft carry case provided.

4. Care and Cleaning

4.1 General Care

All Huntleigh products have been designed to withstand normal clinical use, however they can contain delicate components, for example the probe tip, which should be handled and treated with care.

Periodically, and whenever the integrity of the system is in doubt, carry out a check of all functions as described in the relevant section of the IFU. If there are any defects to the housing contact Huntleigh or your distributor for repair or to order a replacement.



Please ensure that you check with your facility's local infection control policy and medical equipment cleaning procedures.



Observe warnings and guidance on cleaning fluid labelling regarding use and personal protective equipment (PPE).



Do not use abrasive cloths or cleaners.



Do not use automatic washers or autoclaves.



Phenolic detergent based disinfectants, solutions containing cationic surfactants, ammonia based compounds or perfumes and antiseptic solutions such as Steriscol or Hibiscrub should never be used.



If detergent or disinfectant wipes are used ensure that excess solution is squeezed from the wipe prior to use.



Do not allow any fluid to enter the products and do not immerse in any solution.



Always wipe off disinfectant using a cloth dampened with clean water.

4.2 General Cleaning and Disinfecting

Always keep the external surfaces clean and free of dirt and fluids using a clean dry cloth.

1. Wipe any fluids from the surface of the product using a clean dry cloth.
2. Wipe with a cloth dampened in 70% Isopropyl Alcohol.
3. Completely dry with a clean, dry lint free cloth.
4. If the product has been contaminated use the methods described for patient applied parts.

4.3 Cleaning and Disinfecting Patient Applied Parts

Clean the probes before examining a patient using low risk cleaning method below.

Following patient examination, clean and/or disinfect the probes by the appropriate method based upon the level of cross contamination risk, as defined below:

Risk	Definitions	Procedure
Low	Normal use or low risk situations include patients having intact skin and no known infection.	<ol style="list-style-type: none"> 1. Remove soiling, wipe with a mild neutral detergent and then wipe with a cloth dampened in water. 2. Completely dry with a clean lint free cloth.
Medium	The patient has a known infection, skin is not intact, the part is heavily soiled.	<ol style="list-style-type: none"> 1. Follow low risk procedure then wipe with a cloth dampened in Sodium Hypochlorite (1,000ppm). 2. After two minutes wipe with a cloth dampened in water and then dry with a clean lint free cloth.
High	This procedure should only be used when the part has been contaminated by blood.	<ol style="list-style-type: none"> 1. Follow low risk procedure then wipe with a cloth dampened in Sodium Hypochlorite (10,000ppm). 2. After two minutes wipe with a cloth dampened in water and then dry with a clean lint free cloth.



Warning: Sodium Hypochlorite @ 10,000 ppm for disinfecting should only be used in situations described in the High Risk definition. Unnecessary use of this concentrated solution for Low and Medium risk situations may result in damage to the product. Do not allow Sodium Hypochlorite solutions to come into contact with metal parts.

The use of disinfectant materials other than those listed is the responsibility of the user for their efficacy and compatibility with the device.

4.4 Maintenance and Repair

There are NO USER SERVICEABLE PARTS inside the control unit or probe.

Inspection is recommended each time the product is used, paying particular attention to the tip of the probes, checking for cracks etc., and to the cable and connector. Any crackling or intermittent behaviour should be investigated.



This product does not require periodic maintenance.

Suitable test equipment and a full range of spare parts are also available. Please refer to service manual for further information and part numbers.

A full technical description is provided in the Service Manual 726374.

5. Specifications

5.1 Equipment Classification

Type of protection against electric shock.	Internally powered equipment
Degree of protection against electric shock 	Type B - equipment with an applied part. 
Mode of operation.	Continuous
Degree of protection against harmful ingress of particles and/or water.	Main Unit: IP20 Probes (Tip only): IPX1
Degree of safety of application in the presence of a flammable anaesthetic	Equipment not suitable for use in the presence of a FLAMMABLE ANAESTHETIC MIXTURE WITH AIR, OXYGEN OR NITROUS OXIDE

5.2 Standards Compliance

IEC60601-1: 1988 + A1:1991 +A2:1995
UL60601-1 : 2006
CSA C22.2 No 601.1-M90 (R2005)

5.3 FHR Performance*



Standard Mode	Range - 60-210bpm Averaging - 4 beats	Resolution - 1bpm Accuracy - ± 3 bpm
Smoothed Mode	Range - 60-210bpm Averaging - 8 beats	Resolution - 1bpm Accuracy - ± 3 bpm
Manual Mode	Range - 60-210bpm Averaging - 10 beats	Resolution - 1bpm Accuracy - ± 3 bpm

* (excluding user error)

5.4 Waveform Outputs (MD2 Only)

Analogue	Zero crosser, 3.5V full scale per channel (forward and reverse). Conversion factor automatically adjusted to give full scale outputs at $\pm 0.1\%$ ($\pm 10\%$) of probe frequency (e.g. $\pm 8\text{kHz}$ for VP8)
Digital	Data formatted to interface with Printa II or Recorder software running on compatible computer. Conversion factors automatically set as per analogue outputs.
CAL Function	Automatic sequence of CAL levels set to $\pm 0.05\%$ ($\pm 10\%$) and $\pm 0.1\%$ ($\pm 10\%$) of probe frequency (e.g. $\pm 4\text{kHz}$ & $\pm 8\text{kHz}$ for VP8) with zero baseline at start and end of sequence.


5.5 General

Max. Audio Output (Loudspeaker)	500mW rms typical
Auto shut-off	3 minute no signal 10 minute unconditional
Headphones 	Max. output Power: 25 mW rms (32 Ω) Connector: 3.5mm stereo jack socket Max. applied voltage: +9Vdc
RS232 Interface (MD2 only) 	Data format: RS232C Connector: 8pin subminiature DIN socket Max. applied voltage: +5Vdc
Battery Type	IEC 6LR61 or IEC 6LP3146
Battery Life	Typically, 250 x 1 minute examinations
Size	Height 5.5", Depth 1.1", Width 2.9"
Weight	10oz including probe and battery

5.6 Environmental

Operating		Storage
+50°F to +86°F (+10°C to +30°C)	Temperature range	+14°F to +104°F (-10°C to +40°C)
10% to 90% (non condensing)	Relative Humidity	93% maximum
860mb to 1060mb	Pressure	860mb to 1060mb

6. End of Life Disposal



This symbol signifies that this product, including its accessories and consumables is subject to the WEEE (Waste Electrical and Electronic Equipment) regulations and should be disposed of responsibly in accordance with local procedures.

7. Warranty

a) ARJOHUNTLEIGH INC. HEREBY DISCLAIMS ALL EXPRESS OR IMPLIED WARRANTIES (INCLUDING BUT NOT LIMITED TO WARRANTIES OF MERCHANTABILITY OR FITNESS FOR A PARTICULAR PURPOSE) AND ANY AGREEMENTS, REPRESENTATIONS, AFFIRMATIONS, OR WARRANTIES, WHETHER ORAL OR WRITTEN, MADE BY ANY AGENT, EMPLOYEE OR REPRESENTATIVE OF ARJOHUNTLEIGH INC., UNLESS SPECIFICALLY SET FORTH IN THIS PARAGRAPH. ARJOHUNTLEIGH INC. SHALL NOT BE LIABLE FOR BREACH OF CONTRACT ARISING FROM ANY DEFECT IN MATERIAL OR WORKMANSHIP OF THE GOODS. ALL LEGISLATION RELATING TO EXPRESS AND IMPLIED WARRANTIES OR OTHER OBLIGATIONS ON THE PART OF ARJOHUNTLEIGH INC. THAT MAY BE LAWFULLY EXCLUDED ARE HEREBY EXCLUDED.

b) Notwithstanding the foregoing, ArjoHuntleigh Inc.'s sole warranty is that the Goods shall be free from defects in material and workmanship for a period of three (3) years (excluding probe head and retractile cable which are warranted for one (1) year, following delivery of such Goods to the original purchaser; provided that the Goods were used in an appropriate and reasonable manner during such period and provided further that ArjoHuntleigh Inc. shall be in no event be liable to Customer for defective Goods if: (i) the Goods are damaged in the course of shipping; (ii) any defect is caused wholly or to any material extent by customer's negligence, misuse, failure to use the Goods properly or use of the Goods in conjunction with any accessory not approved for use with the Goods by ArjoHuntleigh Inc.; (iii) the Goods are damaged as a result of improper maintenance, failure to follow manufacturer's instructions, including without limitation those on washing and cleaning, or failure to follow necessary routine maintenance procedures; or (iv) the Goods are altered, repaired or dismantled other than with manufacturer's written authorization using its approved procedures or by any party other than manufacturer's properly qualified and trained technicians.

c) Customer must provide written notice to ArjoHuntleigh Inc., within said warranty period of any defect in the Goods. Upon ArjoHuntleigh Inc.'s written request, Customer must return such Goods adequately packed (in their original packing) and fully insured to ArjoHuntleigh Inc.'s place of business and shall be responsible for all shipping costs incurred therein. Customer's exclusive remedy and ArjoHuntleigh Inc.'s exclusive liability for any claim for loss, damage or destruction resulting from any defects in materials and workmanship shall be limited to repair, service, adjustment or replacement (at ArjoHuntleigh Inc.'s option) of any nonconforming or defective Goods. ArjoHuntleigh Inc. will have a reasonable time to repair, service or replace such Goods. Any Goods returned to ArjoHuntleigh Inc. which are found not to be defective in breach of the warranty in Subsection (b) above, shall be returned to the Customer in the manner described in this subsection.

d) IN NO EVENT SHALL ARJOHUNTLEIGH INC. BE LIABLE FOR ANY INCIDENTAL OR CONSEQUENTIAL LOSSES OR DAMAGES (INCLUDING BUT NOT LIMITED TO ECONOMIC LOSS, LOSS OF PROFITS OR SPECIAL DAMAGES) ARISING OUT OF OR INCURRED BY CUSTOMER IN CONNECTION WITH THE PURCHASE OF ARJOHUNTLEIGH INC.'S GOODS EVEN IF ARJOHUNTLEIGH INC. HAS BEEN ADVISED OR HAS KNOWLEDGE OF THE POSSIBILITY OR EXTENT OF SUCH DAMAGES SUFFERED OR INCURRED BY CUSTOMER OR ANY END USER AS A RESULT OF OR IN CONNECTION WITH ANY BREACH OF THESE TERMS AND CONDITIONS BY ARJOHUNTLEIGH INC. OR ANY TORT (INCLUDING BUT NOT LIMITED TO STRICT LIABILITY OR NEGLIGENCE) COMMITTED BY ARJOHUNTLEIGH INC., ITS AGENTS OR REPRESENTATIVES IN CONNECTION WITH THESE TERMS AND CONDITIONS OR ANY CONTRACT WITH CUSTOMER FOR THE SUPPLY OF GOODS.

e) Customer shall not create, directly or indirectly, any warranty obligations on the part of ArjoHuntleigh Inc. to the customers of Customer, and in particular, without limiting the foregoing, Customer agrees not to pass on to its customers any warranties beyond or in addition to those given by ArjoHuntleigh Inc. to Customer hereunder. Where the Customer is a dealer in the Goods, it shall be responsible for the labor cost of all repairs and ArjoHuntleigh Inc. shall be responsible for providing all repair parts during said three (3) year (excluding probe head and retractile cable which are warranted for one (1) year). The dealer shall provide written verification of warranty repairs including the original invoice number, date of purchase, description of repairs, name of its customer and date of sale to such customer.

f) Customer shall be deemed to have full knowledge of the nature and properties of the Goods ordered and of any hazards they involve and the proper treatment, storage and handling thereof. Any technical advice furnished by ArjoHuntleigh Inc. or its representatives or agents is given only on the basis that it is followed at the Customer's own risk.

8. Service

Service Returns

If for any reason Dopplex unit has to be returned, please:

- Clean the product following the instructions in this manual.
- Pack it in suitable packing.
- Attach a decontamination certificate (or other statement declaring that the product has been cleaned) to the outside of the package.
- Call Huntleigh in Addison, IL for a return authorization number.

Huntleigh Healthcare Ltd reserve the right to return product that does not contain a decontamination certificate.

For service, maintenance and any questions regarding this, or any other Huntleigh Healthcare Dopplex product, please contact:

Service Department
ArjoHuntleigh Inc.
2349 West Lake Street,
Addison,
IL 60101, USA

T: 1-800-323-1245 option 2

W: www.ArjoHuntleigh.com

© Huntleigh Healthcare Ltd
All rights reserved



The Dopplex doppler is in conformity with the Medical Devices Directive 93/42/EEC as amended by 2007/47/EC and has been subject to the conformity assurance procedures laid down by the Council Directive

Manufactured in the UK by Huntleigh Healthcare Ltd.
As part of the ongoing development programme the company reserves the right to modify specifications and materials without notice.

Dopplex, and Huntleigh are registered trademarks of Huntleigh Technology Ltd.
2004.

© Huntleigh Healthcare Ltd. 2004-2014

HUNTLEIGH *...performance for life*

Huntleigh Healthcare Ltd. - Diagnostic Products Division
35 Portmanmoor Road, Cardiff, CF24 5HN, United Kingdom
T: +44 (0)29 20485885 F: +44 (0)29 20492520
E: sales@huntleigh-diagnostics.co.uk
W: www.huntleigh-diagnostics.com

DISTRIBUTED IN THE USA BY:

ArjoHuntleigh Inc.
2349 West Lake Street, Addison, IL 60101, USA
T: 800-223-1218
W: www.ArjoHuntleigh.com



Registered No: 942245 England & Wales. Registered Office: ArjoHuntleigh House, Houghton Hall Business Park, Houghton Regis, Bedfordshire, LU5 5XF

© Huntleigh Healthcare Limited 2004-2014

AN **ARJO**HUNTLEIGH COMPANY, MEMBER OF THE GETINGE GROUP

® and ™ are trademarks of Huntleigh Technology Limited

As our policy is one of continuous improvement, we reserve the right to modify designs without prior notice.

726331-US-10
(ENGLISH)